## **Product Data Sheet 546**

# BioMag® Magnetic Immobilization Kit & BioMag® Amine

9025 Technology Dr. • Fishers, IN 46038-2886 800.387.0672 • 317.570.7020 • Fax 317.570.7034 info@bangslabs.com • www.bangslabs.com



## BEADS ABOVE THE REST<sup>™</sup>

#### **DESCRIPTION**

BioMag® Amine consists of a suspension of magnetic iron oxide particles coated to provide primary amine groups. The amine groups allow for the covalent attachment of proteins or ligands with retention of biological activity. BioMag® Amine may be purchased separately or as a component in the BioMag Magnetic Immobilization Kit, which is designed to introduce the user to the ease of magnetizing biological molecules. The BioMag® Magnetic Immobilization Kit includes BioMag® Amine, glutaraldehyde (the reagent used to couple proteins to BioMag®), a BioMag® Separator, and a Reaction Flask. Supplied with the kit is 10mL of BioMag® suspension, sufficient to attach up to 100mg of most proteins. A coupling efficiency of greater than 80% is achieved for many proteins. Higher amounts of protein can be attached at lower coupling efficiencies.

Proteins can be covalently attached to BioMag® Amine by any of the reagents used to prepare affinity supports provided the solid phase terminates with a primary amine group.¹ A suggested glutaraldehyde procedure is given below which follows the method of Weston and Avrameas.²

#### **CHARACTERISTICS**

Mean Diameter: ~1.5μm Particle Concentration: 50 mg/mL

Surface Titration: ~240 µmol/g, ~12µmol/mL

### **MATERIAL**

## Material Supplied in the BioMag® Magnetic Immobilization Kit

- BioMag<sup>®</sup> Amine aqueous suspension: 10mL
- Glutaraldehyde (EM grade\*), 25% solution: 5mL
- Reaction Flask: a single, flat tissue culture vessel
- BioMag<sup>®</sup> Flask Separator

## Material Required for Use with the BioMag® Magnetic Immobilization Kit

- Sodium azide (NaN<sub>a</sub>)
- Pyridine
- Glycine
- EDTA
- Bovine Serum Albumin (BSA)
- Distilled water
- Sodium chloride (NaCl)
- Sodium hydroxide (NaOH)
- Tris base
- Hydrochloric acid (HCl)

#### **PROCEDURE**

Researchers are advised to optimize the use of particles in any application.

Activation and protein coupling steps should be performed in a well-ventilated chemical fume hood.

#### **Preparation of Solutions**

Solution Coupling Buffer	Composition 0.01M pyridine	Materials 0.8mL pyridine	Preparation Instructions Add 0.8mL pyridine to 900mL distilled water. Adjust to pH 6.0 with 6N HCI. Fill to 1L with water.
Glutaraldehyde	5% glutaraldehyde	5mL 25% glutaraldehyde	In a hood, add 5mL glutaraldehyde to 20mL Coupling Buffer.
Glycine Quenching Solution	1M glycine	7.5g glycine	Dissolve 7.5g glycine in 90mL distilled water and adjust to pH 8.0 with 10N NaOH. Fill to 100mL with water.
Wash Buffer	$\begin{array}{l} 0.01 \mathrm{M\ Tris} \\ 0.1 \% \ \mathrm{NaN_3} \\ 0.1 \% \ \mathrm{w/v\ BSA} \\ 0.15 \mathrm{M\ NaCl} \\ 0.001 \mathrm{M\ EDTA} \end{array}$	$\begin{array}{l} \text{1.21g Tris} \\ \text{1.0g NaN}_3 \\ \text{1g BSA} \\ \text{8.7g NaCl} \\ \text{0.37g EDTA} \end{array}$	Dissolve solids in 900mL distilled water. Adjust to pH 7.4 with 10N Na0H or 6N HCl as required. Fill to 1L with water.

#### Activation

- Transfer 10mL of BioMag<sup>®</sup> Amine (equivalent to 500mg of BioMag<sup>®</sup>) to a reaction flask which will easily contain the maximum volume of 50mL used in the coupling procedure. *Note:* A 50mL tissue flask culture of conical tube is typically used.
- 2. Add Coupling Buffer to a final volume of 50mL, shake vigorously, and magnetically separate perpendicular to gravity until the supernatant is clear (approximately 10 minutes). The reaction flask may be secured to the BioMag® Separator with a rubber band. Aspirate the supernatant, leaving the BioMag® as a wet cake on the container wall.
- 3. Repeat Step 2, three times.
- 4. Add 20mL of Glutaraldehyde (5%) to the wet cake and shake vigorously.
- 5. Rotate at room temperature for 3 hours.
- Magnetically separate perpendicular to gravity and aspirate the unreacted glutaraldehyde.
- 7. Repeat Step 2, four times.

#### **Protein Coupling**

 Add 25-100mg of protein to 10mL of Coupling Buffer. Note: For monoclonal antibodies which may be expensive and are supplied at low concentrations (1mg/mL), a carrier protein, such as BSA may be added to increase protein concentration. Various classes of monoclonal antibodies have been successfully coupled to BioMag® by offering 15mg of the monoclonal antibody with 100mg of BSA to 500mg of

Bangs Laboratories, Inc. Product Data Sheet 546 Rev. #011, Active: 02/May/2013 << COPY >> Page 1 of 2

- BioMag® Amine particles. The total volume of the suspension should be about 15mL. Shake vigorously or vortex.
- 2. Remove 75µL of the protein solution and add it to 1mL of Coupling Buffer. Label as Pre-Coupling Solution. Set aside for Coupling Efficiency Determination.
- Add the remaining protein solution to the glutaraldehyde-activated BioMag<sup>®</sup> from Activation, Step 7. Shake vigorously and rotate 16-24 hours at room temperature.
- 4. Magnetically separate and save the supernatant. Label as Post-Coupling Solution. Set aside for Coupling Efficiency Determination.
- Add 50mL of Glycine Quenching Solution and shake vigorously. Rotate for 30 minutes at room temperature.

#### **Washing and Diluting Coupled Particles**

- 1. Magnetically separate and aspirate the supernatant.
- 2. Add 50mL of Wash Buffer and shake vigorously or vortex.
- Magnetically separate perpendicular to gravity, aspirate, and save the supernatant. Add to Post-Coupling Solution and set aside for Coupling Efficiency Determination.
- 4. Repeat Steps 2 and 3, three times.
- 5. Store the coupled BioMag® at 2-8°C as a suspension in Wash Buffer.

#### **Coupling Efficiency Determination**

- Set spectrophotometer wavelength to 280nm. Blank with the Coupling Buffer.
- Measure the absorbance of the Pre-Coupling Solution. A further dilution may be necessary to read an absorbance depending upon the amount of protein added. Recall that the initial dilution made was 75µL in 1mL; a dilution factor (D) of 13.3.
- 3. Measure the absorbance of the Post-Coupling Solution. A dilution may be necessary to read the absorbance.
- 4. Calculate the coupling efficiency, expressed as the % Protein Uptake, as follows. Typical values of Protein Uptake are >60%.

[( $A_{280}$  Pre-Coupling Solution x D) - ( $A_{280}$  Post-Coupling Solution x D)] x 100 ( $A_{280}$  Pre-Coupling Solution x D)

#### **NOTES**

- 1. Phosphate buffer (0.01M, pH 7.0) can be used as a coupling buffer, but with reduced coupling efficiency compared to the recommended pyridine buffer. The polyvalent, negative phosphate ions clump the positively charged amine support. Do not use primary amines, ammonium ion, or other strong nucleophiles in the coupling buffer. All coupling buffers should be used at minimal ionic strengths. Buffers containing amines (e.g. Tris) or phosphate buffers (e.g. PBS) can be used as Wash Buffers. Ionic strength has little or no effect on BioMag® once protein is attached.
- 2. Some noncovalent adsorption invariably accompanies covalent coupling to particulate supports. Noncovalent adsorption is controlled by the washing procedure used after covalent protein attachment. The degree of noncovalent adsorption varies with each application and the washing procedure may need to be adjusted for individual applications. Additional washes to reduce noncovalently adsorbed protein can include high salt (1M NaCl), mildly acidic or basic media, mildly elevated temperatures, or increased time of exposure to Wash Buffer. Dissociation of active, noncovalently adsorbed molecules from BioMag® can make magnetic materials appear unstable in some applications.
- 3. Prolonged vigorous shaking or vortexing should be used to resuspend

BioMag® after magnetic separation or settling with gravity.

#### REFERENCES

- 1. **Avrameas, S.** 1969. Coupling of enzymes to proteins with glutaraldehyde. Use of conjugates for the detection of antigens and antibodies. *Immunochemistry*, 6(1): 43-52.
- Weston, P.D., S. Avrameas. 1971. Proteins coupled to polyacrylamide beads using glutaraldehyde. *Biochem Biophys Res Commun*, 45(6): 1574-1580.

#### STORAGE AND STABILITY

Store at 2-8°C. Freezing, drying, or centrifuging BioMag® may result in irreversible aggregation and loss of binding activity.

These products are for research use only and are not intended for use in humans or for *in vitro* diagnostic use.

#### **ORDERING INFORMATION**

Cat. Code	Description	Sizes
BM545	BioMag® Magnetic Immobilization Kit	1 kit
BM546	BioMag® Amine	10mL or 100mL

Order online anytime at www.bangslabs.com.

Bangs Laboratories, Inc. Product Data Sheet 546 Rev. #011, Active: 02/May/2013 << COPY >> Page 2 of 2