

9025 Technology Dr. • Fishers, IN 46038-2886
 800.387.0672 • 317.570.7020 • Fax 317.570.7034
 info@bangslabs.com • www.bangslabs.com



B E A D S • A B O V E T H E R E S T™

DESCRIPTION

The CD2 antigen, also known as Leukocyte Function-associated Antigen 2 (LFA-2), is a monomeric 50 kDa glycoprotein that is a receptor for sheep erythrocytes and the CD58/LFA-3 ligand. CD2 plays a role in T cell signaling and lymphocyte adhesion. It is expressed on the vast majority of thymocytes and all peripheral T lymphocytes, and is also expressed on malignant cells of T origin including T ALL and a subset of NK cells.

BioMag® anti-Human CD2 particles are designed for positive selection of CD2 plus human T cells.

BioMag® anti-Human CD2 is a suspension of magnetic particles approximately 1.5µm in size. The suspension is supplied in a phosphate buffered saline (pH 7.5) containing EDTA, 1.0% BSA, and 0.1% sodium azide.

CHARACTERISTICS

Mean Diameter: ~1.5µm
 Particle Concentration: 4 mg/mL
 Particle Count: 1 x 10⁸ BioMag® particles per mg

PROCEDURE

Researchers are advised to optimize the use of particles in any application.


Depending on antigen availability and the size of the target cell population, cell sorting applications may require up to 50-60 magnetic particles per cell. Magnetic particles and cells should be incubated at room temperature for 30-60 minutes in media containing 5-10% protein (to reduce nonspecific binding) for successful separation. Gentle end-over-end mixing or rocking during incubation is required for optimal results. (*Note: Increasing the incubation time beyond one hour may be necessary to achieve the desired depletion.*)

Some applications require the detachment of BioMag® antibody particles from cells after separation. One approach would involve culturing cells after positive selection. Cultures can be maintained for about 48 hours during which magnetic particles fall away from cells due to cell surface changeover. The magnetic particles are then easily removed via a magnetic separation. Another approach is the use of a protease, such as chymopapain, to break the antigen-antibody bond and remove the particles magnetically. Depending upon the application, it may not be necessary to remove the cells from the BioMag® particles. BioMag® particles are only 1-2µm in size and have been successfully used in FACS equipment. They will not jam the machine and are distinguishable from cells. Alternatively, negative selection approaches can be very effective in producing specific cell populations.

STORAGE AND STABILITY

Store at 2-8°C. Freezing, drying, or centrifuging BioMag® may result in irreversible aggregation and loss of binding activity. Washing BioMag® anti-

Cell sorting results using BioMag® anti-Human CD2 leukocyte particles for positive selection. Typically, whole blood or purified leukocytes and particles are incubated for 30 minutes at room temperature and then magnetically separated. The supernatant is collected, incubated with the appropriate two-color antibody cocktail, and then analyzed by flow cytometry. Figure A depicts the cell population prior to positive selection. Figure B depicts the cell population after positive selection. The particle to cell ratios reported above are based on experiments where cells were exposed to the particles once.



General Recommendation*:

| | |
|----------------------------|-------------------------------------|
| Concentration# | 4.00 x 10 ⁸ particles/mL |
| Volume Used | 50µL |
| # Particles | 2 x 10 ⁷ per test |
| # Target Cells | ~7.32 x 10 ⁶ per mL |
| Particle:Target Cell Ratio | 27:1 |
| % Depletion | 94.0% |

** These values should be used as a starting point in optimizing experimental protocols. Due to differences in the distribution of cell types in samples and other variables, the researcher is strongly encouraged to determine the optimal particle to cell ratios for their experiments.*

Human CD2 particles in sterile media to remove preservative prior to use is recommended. Using a magnetic separation unit for washing instead of centrifugation is also strongly recommended.

SAFETY

This particle suspension contains sodium azide. Sodium azide may react with lead and copper plumbing to form explosive metal azides. Upon disposal of material, flush with a large volume of water to prevent azide accumulation. Please consult the Material Safety Data Sheet for more information.

These products are for research use only and are not intended for use in humans or for *in vitro* diagnostic use.

ORDERING INFORMATION

| Cat. Code | Description | Size |
|-----------|------------------------|------|
| BM595 | BioMag® anti-Human CD2 | 5mL |

Order online anytime at www.bangslabs.com.